



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

FAIRBANKS SCALES INC.
23471 Lakepointe Drive
Clinton Township, MI 48036-4801
Joe Ehrnschwender Phone: 513 860 8062

CALIBRATION

Valid To: December 31, 2025

Certificate Number: 1843.02

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations^{1, 4}:

I. Mechanical

| Parameter/Equipment | Range | CMC ² (±) | Comments |
|-----------------------|------------------------|----------------------|---------------------------|
| Balances ³ | (0 to 100) g | 0.30 mg | Class 1, 6, and F weights |
| | (100 to 500) g | 2.2 mg | |
| | (500 to 4000) g | 0.11 g | |
| | (4000 to 25 000) g | 3.0 g | |
| Scales ³ | (0 to 100) lb | 0.011 lb | Class 6 and F weights |
| | (101 to 500) lb | 0.035 lb | |
| | (501 to 1000) lb | 0.12 lb | |
| | (1001 to 5000) lb | 0.39 lb | |
| | (5001 to 10 000) lb | 0.69 lb | |
| | (10 001 to 200 000) lb | 50 lb | |

¹ This laboratory offers commercial and field calibration service.

- ² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.
- ³ Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g., resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.
- ⁴ This scope meets A2LA's *P112 Flexible Scope Policy*.