



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

FAIRBANKS SCALES INC.  
23471 Lakepointe Drive  
Clinton Township, MI 48036-4801  
Bob Raymond Phone: 513 860 8062

CALIBRATION

Valid To: December 31, 2027

Certificate Number: 1843.02

In recognition of the successful completion of the A2LA evaluation process (including an assessment of the organization's compliance with R205 – A2LA's Calibration Program Requirements), accreditation is granted to this laboratory to perform the following calibrations<sup>1,4</sup>:

I. Mechanical

| Parameter/Equipment   | Range                  | CMC <sup>2</sup> (±) | Comments                  |
|-----------------------|------------------------|----------------------|---------------------------|
| Balances <sup>3</sup> | (0 to 100) g           | 0.30 mg              | Class 1, 6, and F weights |
|                       | (100 to 500) g         | 2.2 mg               |                           |
|                       | (500 to 4000) g        | 0.11 g               |                           |
|                       | (4000 to 25 000) g     | 3.0 g                |                           |
| Scales <sup>3</sup>   | (0 to 100) lb          | 0.011 lb             | Class 6 and F weights     |
|                       | (101 to 500) lb        | 0.035 lb             |                           |
|                       | (501 to 1000) lb       | 0.12 lb              |                           |
|                       | (1001 to 5000) lb      | 0.39 lb              |                           |
|                       | (5001 to 10 000) lb    | 0.69 lb              |                           |
|                       | (10 001 to 200 000) lb | 50 lb                |                           |

<sup>1</sup> This laboratory offers commercial and field calibration service.

- <sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.
- <sup>3</sup> Field calibration service is available for this calibration. Please note the actual measurement uncertainties achievable on a customer's site can normally be expected to be larger than the CMC found on the A2LA Scope. Allowance must be made for aspects such as the environment at the place of calibration and for other possible adverse effects such as those caused by transportation of the calibration equipment. The usual allowance for the actual uncertainty introduced by the item being calibrated, (e.g., resolution) must also be considered and this, on its own, could result in the actual measurement uncertainty achievable on a customer's site being larger than the CMC.
- <sup>4</sup> This scope meets A2LA's *P112 Flexible Scope Policy*.



# Accredited Laboratory

A2LA has accredited

## FAIRBANKS SCALES INC.

*Clinton Township, MI*

for technical competence in the field of

## Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated April 2017*).



Presented this 25<sup>th</sup> day of October 2025.

A blue ink signature of Mr. Trace McInturff.

Mr. Trace McInturff, Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 1843.02  
Valid to December 31, 2027

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*